

PATENT COOPERATION TREATY

PCT

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY (Chapter I of the Patent Cooperation Treaty)

(PCT Rule 44bis)

Applicant's or agent's file reference P36489-P0	FOR FURTHER ACTION	See item 4 below
International application No. PCT/JP2004/019508	International filing date (<i>day/month/year</i>) 27 December 2004 (27.12.2004)	Priority date (<i>day/month/year</i>) 26 December 2003 (26.12.2003)
International Patent Classification (8th edition unless older edition indicated) See relevant information in Form PCT/ISA/237		
Applicant MATSUSHITA ELECTRIC INDUSTRIAL CO., LTD.		

1. This international preliminary report on patentability (Chapter I) is issued by the International Bureau on behalf of the International Searching Authority under Rule 44 bis.1(a).
2. This REPORT consists of a total of 4 sheets, including this cover sheet.

In the attached sheets, any reference to the written opinion of the International Searching Authority should be read as a reference to the international preliminary report on patentability (Chapter I) instead.

3. This report contains indications relating to the following items:

- | | | |
|-------------------------------------|--------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <input checked="" type="checkbox"/> | Box No. I | Basis of the report |
| <input type="checkbox"/> | Box No. II | Priority |
| <input type="checkbox"/> | Box No. III | Non-establishment of opinion with regard to novelty, inventive step and industrial applicability |
| <input type="checkbox"/> | Box No. IV | Lack of unity of invention |
| <input checked="" type="checkbox"/> | Box No. V | Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement |
| <input type="checkbox"/> | Box No. VI | Certain documents cited |
| <input type="checkbox"/> | Box No. VII | Certain defects in the international application |
| <input type="checkbox"/> | Box No. VIII | Certain observations on the international application |

4. The International Bureau will communicate this report to designated Offices in accordance with Rules 44bis.3(c) and 93bis.1 but not, except where the applicant makes an express request under Article 23(2), before the expiration of 30 months from the priority date (Rule 44bis .2).

BEST AVAILABLE COPY

The International Bureau of WIPO 34, chemin des Colombettes 1211 Geneva 20, Switzerland Facsimile No. +41 22 338 82 70	Date of issuance of this report 22 August 2006 (22.08.2006) Authorized officer <p style="text-align: center;">Yoshiko Kuwahara</p> e-mail: pt07@wipo.int
-------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------------------------------------

PATENT COOPERATION TREATY

TRANSLATION

From the
INTERNATIONAL SEARCHING AUTHORITY

PCT

WRITTEN OPINION OF THE
INTERNATIONAL SEARCHING AUTHORITY

(PCT Rule 43bis.1)

To:

Date of mailing
(day/month/year)

Applicant's or agent's file reference

P36489-P0

FOR FURTHER ACTION

See paragraph 2 below

International application No.

PCT/JP2004/019508

International filing date (day/month/year)

27.12.2004

Priority date (day/month/year)

26.12.2003

International Patent Classification (IPC) or both national classification and IPC

Applicant

MATSUSHITA ELECTRIC INDUSTRIAL CO., LTD.

1. This opinion contains indications relating to the following items:

- ☒ Box No. I Basis of the opinion
- ☐ Box No. II Priority
- ☐ Box No. III Non-establishment of opinion with regard to novelty, inventive step and industrial applicability
- ☐ Box No. IV Lack of unity of invention
- ☒ Box No. V Reasoned statement under Rule 43bis.1(a)(i) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement
- ☐ Box No. VI Certain documents cited
- ☐ Box No. VII Certain defects in the international application
- ☐ Box No. VIII Certain observations on the international application

2. **FURTHER ACTION**

If a demand for international preliminary examination is made, this opinion will be considered to be a written opinion of the International Preliminary Examining Authority ("IPEA") except that this does not apply where the applicant chooses an Authority other than this one to be the IPEA and the chosen IPEA has notified the International Bureau under Rule 66.1bis(b) that written opinions of this International Searching Authority will not be so considered.

If this opinion is, as provided above, considered to be a written opinion of the IPEA, the applicant is invited to submit to the IPEA a written reply together, where appropriate, with amendments, before the expiration of 3 months from the date of mailing of Form PCT/ISA/220 or before the expiration of 22 months from the priority date, whichever expires later.

For further options, see Form PCT/ISA/220.

3. For further details, see notes to Form PCT/ISA/220.

Name and mailing address of the ISA/JP

Authorized officer

Facsimile No.

Telephone No.

WRITTEN OPINION OF THE
INTERNATIONAL SEARCHING AUTHORITY

International application No.

PCT/JP2004/019508

Box No. I Basis of this opinion

1. With regard to the language, this opinion has been established on the basis of the international application in the language in which it was filed, unless otherwise indicated under this item.

☐

This opinion has been established on the basis of a translation from the original language into the following language

_____, which is the language of a translation furnished for the purposes of international search (under Rule 12.3 and 23.1(b)).

2. With regard to any nucleotide and/or amino acid sequence disclosed in the international application and necessary to the claimed invention, this opinion has been established on the basis of:

a. type of material

☐

a sequence listing

☐

table(s) related to the sequence listing

b. format of material

☐

in written format

☐

in computer readable form

c. time of filing/furnishing

☐

contained in the international application as filed.

☐

filed together with the international application in computer readable form.

☐

furnished subsequently to this Authority for the purposes of search.

3. ☐ In addition, in the case that more than one version or copy of a sequence listing and/or table(s) relating thereto has been filed or furnished, the required statements that the information in the subsequent or additional copies is identical to that in the application as filed or does not go beyond the application as filed, as appropriate, were furnished.

4. Additional comments:

WRITTEN OPINION OF THE
INTERNATIONAL SEARCHING AUTHORITY

International application No.

PCT/JP2004/019508

Box No. V Reasoned statement under Rule 43bis.1(a)(i) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement

1. Statement

Novelty (N)

Claims 1-55 YES

Claims NO

Inventive step (IS)

Claims 2-17, 21, 22, 24, 25, 33-44, 47 YES

Claims 1, 18-20, 23, 26-32, 45, 46, 48-55 NO

Industrial applicability (IA)

Claims 1-55 YES

Claims NO

2. Citations and explanations:

Document 1: JP 2003-504637 A (P E Corporation (NY)), 04 February 2003

Document 2: JP 6-66769 A (Hitachi, Ltd.), 11 March 1994

Claims 1, 23 and 26 described in an independent form includes an invention described in document 1 or 2.

Also, the inventions of claims 18-20, 27-32, 45, 46 and 48-55 described in a cited form are, to a party skilled in the art, empirically technical common sense, so well-known commonly used matters that there is no need to provide examples, or matter of design variation that could be easily conceived of therefrom.

The inventions of claims 2, 3, 24, 25 and 33 have a technical feature such that a plate is rotated to cause a biological sample to flow in to a first flow position not reaching a quantifying unit in a second flow path, and a plate is pressurized to cause the biological sample to flow from the first flow position to a second flow position including the quantifying unit in the second flow path, and such technical feature is neither described nor suggested in any of the documents cited in the ISR.

Therefore, the inventions of claims 2, 3, 24, 25 and 33 could not be easily conceived of by a party skilled in the art.

Thus, since claims 4-17, 21, 22, 34-44 and 47 that cite from one of claims 2, 3, 24, 25 and 33 includes the above technical feature, the inventions thereof could not be easily conceived of by a party skilled in the art.

BEST AVAILABLE COPY

PCL XL error

Subsystem: KERNEL

Error: IllegalOperatorSequence

Operator: 0x1b

Position: 1240

BEST AVAILABLE COPY